Commercial product name / designation Application / Use SIC Use Category Manufacturer Pulpdent Corporation 80 Oakland Street, P.O. Box 780 Watertown, MA 02472 USA Emergency Telephone Number Authorized European Representative UK Responsible Person	851 Human health ac 55 Telephone: 1 617 926 Fax: 1 617 926 Email: Pulpdent@pulp 1-800-535-5053 (24 H Advena Limited Tower Business Cent Tower Street,	e by dental professional onl tivity 6-6666 6-6262 pdent.com Hour / USA) re, 2nd Floor,	ly.		
SIC Use Category Manufacturer Pulpdent Corporation 80 Oakland Street, P.O. Box 780 Watertown, MA 02472 USA Emergency Telephone Number Authorized European Representative	Telephone: 1 617 926 Fax: 1 617 926 Email: Pulpdent@pulp 1-800-535-5053 (24 F Advena Limited Tower Business Cent Tower Street, Swatar, BKR 4013 Ma	5-6666 5-6262 pdent.com Hour / USA) re, 2nd Floor,	ly.		
Manufacturer Pulpdent Corporation 80 Oakland Street, P.O. Box 780 Watertown, MA 02472 USA Emergency Telephone Number Authorized European Representative	Telephone: 1 617 926 Fax: 1 617 926 Email: Pulpdent@pulp 1-800-535-5053 (24 H Advena Limited Tower Business Cent Tower Street, Swatar, BKR 4013 Ma	S-6262 pdent.com Hour / USA) re, 2nd Floor,			
Authorized European Representative	1-800-535-5053 (24 F Advena Limited Tower Business Cent Tower Street, Swatar, BKR 4013 Ma	Hour / USA)			
Authorized European Representative	Advena Limited Tower Business Cent Tower Street, Swatar, BKR 4013 Ma	re, 2nd Floor,			
UK Responsible Person	Advena Limited		Advena Limited Tower Business Centre, 2nd Floor,		
	Pure Offices, Plato Cl Warwick, CV34 6WE				
CH Authorized Representative	MedEnvoy Switzerlan Gotthardstrasse 28, 6				
Hazards Identification					
Classification	Irritant.				
Classification according to Regulation (EC) No 1272/2008 [CLP]	Hazard Class Eye irritation STOT SE Skin irritation Skin sensitization	Hazard Category 2 3 2 1	Hazard Statement H319 H335 H315 H317		
Classification according to Directive 67/548/EEC (See SECTION 16 for full text of risk phrases)	Irritant; Xi; R 36/37/38 - 43				
Hazard Pictograms Signal Word: WARNING Restricted to use by dental professional only. Hazard Statements H319: Eye irritation. 2. May cause eye irritation.					
	See SECTION 16 for full text of risk phrases) GHS Label Elements dazard Pictograms Gignal Word: WARNING Restricted to use by dental professional only. dazard Statements days: STOT SE. 3. May cause eye irritation. days: STOT SE. 3. May cause respiratory irritation. days: Skin irritation. 2. May cause skin irritation. days: Sensitization. 1. May cause an allergic skin irrecautionary Statements	See SECTION 16 for full text of risk phrases) GHS Label Elements Idazard Pictograms Gignal Word: WARNING Restricted to use by dental professional only. Idazard Statements Iday Statements	See SECTION 16 for full text of risk phrases) SHS Label Elements Idazard Pictograms Signal Word: WARNING Restricted to use by dental professional only. Idazard Statements Iday: Eye irritation. 2. May cause eye irritation. Iday: STOT SE. 3. May cause respiratory irritation. Iday: Skin irritation. 2. May cause skin irritation. Iday: Sensitization. 1. May cause an allergic skin reaction.		

Trade Name: $PULPDENT PRONTO^{TM}$

P280: Wear protective gloves and eye protection

P305+P351: If in eyes, rinse cautiously with water for several minutes.

P337+P313: If eye irritation persists, get medical advice/attention.

P302+P352: If on skin, wash with plenty of soap and water.

P333+P313: If irritation or rash occurs, get medical advice/attention.

P410+P411: Protect from sunlight. Store at temperature not exceeding 27°C / 80°F.

	P410+P411: Protect from sunlight. Store at temperature not exceeding 27°C / 80°F.				
3.0	Composition				
3.1	Chemical charac	cterization of the preparation:	Light-cu	ured, glass-filled resins.	
3.2	Hazardous ingredients				
	CAS Number	Name of the Ingredient	Concentration	Classification per 67/548/EEC	Classification per Regulation (EC) No.1272/2008 (CLP)
	Proprietary mixture	Blend of diurethane and other methacrylate resins	35.0%	Irritant; Xi R 3637//38 - 43	Eye irritation, 2, H319 STOT SE 3, H335 Skin irritation, 2, H315 Skin sensitization, 1, H317
	112945-52-5	Silica, amorphous	4.8%	Irritant; Xi, R 36/37/38	Eye irritation, 2, H319 STOT SE 3, H335 Skin irritation, 2, H315
4.0	First Aid Measur	es			
4.1	General Information		May be irritating to eyes, respiratory system, skin. Prolonged or repeated contact with resins may cause sensitization. Show this safety data sheet to medical personnel. Get medical attention in case of uncertainty.		
4.2	Inhalation		Move to fresh air. If necessary, administer oxygen and/or artificial respiration and seek medical attention.		
4.3	Skin Contact		Take off contaminated clothing. Wash skin thoroughly with soap and water for 15 minutes.		
4.4	Eye Contact		Keep eyelids apart and flush with running water for 15+ minutes. Get medical attention.		
4.5	Ingestion			irritating to mucous mer	vomiting. Get immediate medical mbranes. Never give anything by
4.6	Precautions for first responders		Ventilate the area. Wear eye and skin protection.		
4.7	Information for ph	ysicians			
	Symptoms		Irritation or redness in eyes, throat or on skin.		
	Hazards		May be irritating sensitization by s		system and skin. May cause
	Treatment		As above under F	First Aid.	
5.0	Fire Fighting Me	asures			
5.1	Suitable extinguis	hing media		lry chemical, alcohol foan ire exposed containers c	n, or water fog. Water spray may ool.
5.2	Extinguishing media to avoid		Do not use direct water stream		

8.2.1.5

8.2.2

9.2.1

рΗ

Other controls

Environmental exposure controls

Trade Name: **PULPDENT PRONTO™** 5.3 Special exposure hazards in a fire Heat may cause polymerization with rapid release of energy. 5.4 Special protective equipment for fire-fighters A self-contained breathing apparatus should be worn by firefighters. 6.0 **Accidental Release Measures** 6.1 Personal precautions. Ventilate area. Wear safety glasses, gloves and lab coat. 6.2 Environmental precautions Contain spilled material. Follow all government regulations. 6.3 Method for clean up Absorb or wipe up spill with paper towels or cloths. Collect for disposal in a covered container. Wash area of spill with alcohol or soap and water. 7.0 Handling and Storage 7.1 For use only by dental professionals. Follow good hygiene practices. Handling Remove applicator tip from syringe and recap immediately after use. Keep light-cured materials shaded from intense light sources. 7.2 Keep tightly capped. Store in original container at cool room temperature Storage (<25°C). Avoid direct light, sources of ignition, extremes of temperature (>27°C/80°F, <5°C/40°F). Shelf life for unopened product is two years from date of manufacture, provided that the material has been stored properly. 7.3 Specific uses Dental material 8.0 Exposure Controls / Personal Protection 8.1 Exposure limit values PEL: Not established. TLV: Not established. 8.2 Exposure controls 8.2.1 Occupational exposure controls No special equipment required under normal conditions of use. 8.2.1.1 Respiratory protection Good general ventilation is sufficient to control any airborne vapors. 8.2.1.2 Hand protection Usual surgical gloves will limit contact with uncured material. 8.2.1.3 Eye protection Usual safety glasses. 8.2.1.4 Good personal hygiene and safety practices; lab coat. Skin protection

		is inert. Follow all government regulations.
9.0	Physical and Chemical Properties	
9.1	Characteristics	
9.1.1	Appearance /Color / Physical state	Paste in tooth shades
9.1.2	Odor	Faint, characteristic
9.2	Important health, safety and environmental information	

Not determined

Emergency eye wash fountain should be available. Wash hands after use.

Cure material before disposing. To the best of our knowledge, cured material

9.2.2 Boiling point Not determined
9.2.3 Flash point Not determined
9.2.4 Flammability (solid, gas) Not applicable
9.2.5 Explosive properties Not applicable

14.0	Transport Information	
13.1	Regulations	Polymerize before disposal. Follow all local and national government regulations in disposing material or contaminated packaging.
13.0	Disposal Considerations	
12.1	Ecotoxicity	To the best of our knowledge, polymerized material is inert. No other information is available.
12.0	Ecological Information	
11.7	Clinical experience	Product is new to the market, but initial evaluations have been positive. There have been no reports of adverse events.
11.6	Empirical data	None available.
11.5	Carcinogenicity, Mutagenicity, Reproductive Toxicity	None known.
11.4	Sub-acute, sub-chronic and prolonged toxicity	Prolonged and/or frequent skin contact may cause allergic skin reactions in susceptible individuals. Prolonged exposure to large amounts of this material may cause eye and respiratory irritation.
11.3	Sensitization	May be sensitizing. Prolonged or frequent skin contact may cause allergic skin reactions in some susceptible individuals.
11.2	Irritation and corrosiveness	May be irritating to eyes, mucous membranes or skin on contact or with prolonged exposure.
11.1	Acute toxicity	Not toxic. Minimal health hazard in the quantities present in this product and under normal conditions of use.
11.0	Toxicological information	
10.4	Further information	Polymerization will occur when light-cured material is exposed to direct light.
10.3	Hazardous decomposition products	Under fire conditions, with amounts greater than in this product, hazardous polymerization may occur with heat buildup, release of carbon monoxide, carbon dioxide, oxides of nitrogen.
10.2	Materials to avoid	Reducing and oxidizing agents, peroxides, amines.
10.1	Conditions to avoid	Temperature extremes (>27°C/80°F, <5°C/40°F), light.
10.0	Stability and reactivity	
9.2.13	Evaporation rate	Not determined
9.2.12	Vapor density	>1
9.2.11	Viscosity	Not determined
9.2.10	Partition coefficient	Not determined
9.2.9	Solubility in water	Nil
9.2.8	Specific gravity	1.700
9.2.7	Vapor pressure	< 1 mm Hg / 133 Pa / Id: B
9.2.6	Oxidizing properties	Not determined

Revision Date: June 14, 2022

Safety Data Sheet

14.1	Restrictions None. Not regulated by IATA.	
15.0	Regulatory Information	
15.1	US FDA Class II medical device	

16.0	Other information	
16.1	List of relevant R phrases	R36/37/38: Irritating to eyes, respiratory system and skin. R43: Sensitizing by skin contact
16.2	Hazard Statements	H319: Eye irritation. Hazard category 2. H335: Specific Target Organ Toxicity - Single exposure; hazard category. 3. Respiratory tract irritation. H315: Skin irritation. Hazard category 2. H317: Skin Sensitization. Hazard category 1.
16.3	Precautionary Statements	P261: Avoid breathing vapor. P280: Wear protective gloves and eye protection P305+P351: If in eyes, rinse cautiously with water for several minutes. P337+P313: If eye irritation persists, get medical advice/attention. P302+P352: If on skin, wash with plenty of soap and water. P333+P313: If irritation or rash occurs, get medical advice/attention. P410+P411: Protect from sunlight. Store at temperature not exceeding 27°C / 80°F.
16.4	Restrictions on use	Pulpdent products are to be sold to and used by dental professionals only.
16.5	Further information	The information presented herein is believed to be factual as it has been derived from the works of persons believed to be qualified experts. However, nothing contained in this information is to be taken as a warranty or representation for which Pulpdent Corporation bears legal responsibility. The user should review any recommendations in the specific context of the intended use to determine whether they are appropriate.
16.6	Sources of key data	National Institute for Occupational Safety (NIOSH) US Occupational Safety and Health Administration (OSHA) Eur-Lex European Union Law: Regulation (EC) No. 1272/2008 (CLP) and Regulation (EC) No. 1907/2006 (REACH). Guidance on the compilation of safety data sheets. Version 1.1; December 2011. European Chemicals Agency
16.7	Information which has been added, deleted or revised.	This Safety Data Sheet has been revised to meet the requirements of the GHS SDS format, Regulations (EC) No. 1272/2008 (CLP) and (EC) No. 1907/2006 (REACH). Specifically, Sections 2.1, 2.2, 3.2, 16.2, 16.3 have been modified.